



Product Overview

PRODUCT:	Serum Multiple Calibrator
PART NUMBER:	5589
FORM / APPEARANCE:	Liquid, amber to straw, free of visible foreign matter, clots or precipitates
FILTRATION:	0.22 micron-filtered
ADDITIVES:	None
STORAGE:	<-20 °C
EXPIRATION:	12 months from date of manufacture

ANALYTE	
CREATININE:	4.5-505 mg/dL
UREA NITROGEN:	27-33 mg/dL
URIC ACID:	9-11 mg/dL
PHOSPHOROUS:	9-11 mg/dL
CALCIUM:	9-11 mg/dL
MAGNESIUM:	2.7-3.3 mg/dL
GLUCOSE:	180-220 mg/dL
TOTAL PROTEIN:	6.7-7.4 g/dL
ALBUMIN:	4.8-5.3 g/dL
SODIUM:	130-180 mmol/L
POTASSIUM:	4.0-5.0 mmol/L
CHLORIDE:	95-125 mmol/L
TRIGLYCERIDE:	<10 mg/dL
CHOLESTEROL:	<10 mg/dL
pH:	7.7-7.9
TURBIDITY (O.D. @ 546nm):	< 0.24

FOR IN VITRO DIAGNOSTIC MANUFACTURING USE ONLY

All human plasma units used to prepare this product have been tested by FDA approved tests and found to be non-reactive for HbsAg, Anti-HIV I/II, Anti-HCV, HIV-1 RNA, and HCV RNA. Donors of human plasma units used in the making of this product were tested and found negative for Syphilis according to FDA guidelines.

CAUTION: FOR USE IN RESEARCH OR REMANUFACTURE OF NON-INJECTABLE PRODUCTS NOT SUBJECT TO LICENSE UNDER PHS ACT SECTION 351, ONLY.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE HANDLED AND TREATED AS POTENTIALLY INFECTIOUS