



## Product Overview

PRODUCT:	UIBC Calibrator
PART NUMBER:	4320
FORM / APPEARANCE:	Off-white crystalline powder
FILTRATION:	0.22 micron-filtered
ADDITIVES:	None
STORAGE:	2 – 8 °C
EXPIRATION:	3 years from date of manufacture

## Analytical Properties

ANALYTE	VALUE
TOTAL PROTEIN:	4 – 8 g/dL
UIBC:	235 – 285 µg/dL
SODIUM:	130 – 148 mmol/L
CHLORIDE:	90 – 110 mmol/L
pH:	7.0 – 7.8
MOISTURE:	< 2 %
BIOBURDEN:	< 100 CFUs/mL

### **FOR IN VITRO DIAGNOSTIC MANUFACTURING USE ONLY**

All human plasma units used to prepare this product have been tested by FDA approved tests and found to be non-reactive for HbsAg, Anti-HIV I/II, Anti-HCV, HIV-1 RNA, and HCV RNA. Donors of human plasma units used in the making of this product were tested and found negative for Syphilis according to FDA guidelines.

CAUTION: FOR USE IN RESEARCH OR REMANUFACTURE OF NON-INJECTABLE PRODUCTS NOT SUBJECT TO LICENSE UNDER PHS ACT SECTION 351, ONLY.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE HANDLED AND TREATED AS POTENTIALLY INFECTIOUS.