



Product Overview

PRODUCT:	UIBC Calibrator, Human Based
PART NUMBER:	4554
FORM / APPEARANCE:	Off-white crystalline
FILTRATION:	0.22 micron-filtered
ADDITIVES:	None
STORAGE:	2-8 °C
EXPIRATION:	3 years from date of manufacture

ANALYTE	
UIBC:	235-285
SODIUM:	130-148
CHLORIDE:	90-110
TOTAL PROTEIN:	4-8
pH:	7.0-7.8
VIAL-VIAL VARIATION	<1% CV, Sodium
MOISTURE:	<2%; Karl Fisher
BIOBURDEN:	<100
RECONSTITUTED STABILITY:	14 days, 2-8°C
ACCELERATED STABILITY:	10.5 days, 37°C

FOR IN VITRO DIAGNOSTIC MANUFACTURING USE ONLY

All human plasma units used to prepare this product have been tested by FDA approved tests and found to be non-reactive for HbsAg, Anti-HIV I/II, Anti-HCV, HIV-1 RNA, and HCV RNA. Donors of human plasma units used in the making of this product were tested and found negative for Syphilis according to FDA guidelines.

CAUTION: FOR USE IN RESEARCH OR REMANUFACTURE OF NON-INJECTABLE PRODUCTS NOT SUBJECT TO LICENSE UNDER PHS ACT SECTION 351, ONLY.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE HANDLED AND TREATED AS POTENTIALLY INFECTIOUS