



Product Overview

| | |
|--------------------|---|
| PRODUCT: | hDNA Interference Control |
| PART NUMBER: | 5105 |
| FORM / APPEARANCE: | Liquid, amber to straw, free of visible foreign matter, clots or precipitates |
| FILTRATION: | 0.22 micron-filtered |
| ADDITIVES: | None |
| STORAGE: | < -20 °C |
| EXPIRATION: | 2 years from date of manufacture |

Analytical Properties

| ANALYTE | VALUE |
|---------|------------|
| hDNA: | 2 – 4 mg/L |

FOR IN VITRO DIAGNOSTIC MANUFACTURING USE ONLY

All human plasma units used to prepare this product have been tested by FDA approved tests and found to be non-reactive for HbsAg, Anti-HIV I/II, Anti-HCV, HIV-1 RNA, and HCV RNA. Donors of human plasma units used in the making of this product were tested and found negative for Syphilis according to FDA guidelines.

CAUTION: FOR USE IN RESEARCH OR REMANUFACTURE OF NON-INJECTABLE PRODUCTS NOT SUBJECT TO LICENSE UNDER PHS ACT SECTION 351, ONLY.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE HANDLED AND TREATED AS POTENTIALLY INFECTIOUS